


Declaration of Conformity

EC Declaration of Conformity
under the Medical Devices
Directive 93/42/EEC.



Intelligent Orthopaedics

PO Box 2681
Stoke on Trent
UK ST4 9BE

T +44 (0) 17 82 84 78 40

www.intelligent-orthopaedics.co.uk

Product: Bone Twist Drill
Part Number: SITD - DDDLLLSSS
Details of product:

Single use bone twist drill for use with Intelligent
Orthopaedics Fracture Management Devices (t < 60 minutes)

DDD = diameter

LLL = overall length

SSS = drill flute length

EC Product Class Analysis:

Annex IX: Rule 6

Transient Use surgically invasive device

Class IIa

Notified Body

Amtac Medica

Technical Files are held at Intelligent Orthopaedics Ltd.

A Complaints and Vigilance Procedure is in action and details are supplied
with the purchase contract, they may also be obtained from <http://www.intelligent-orthopaedics.co.uk>.

Approved by: 

Date: 4/1/06.

Dr C Ian Moorcroft
Technical Director
Intelligent Orthopaedics Ltd.

Declaration of Conformity

EC Declaration of Conformity
under the Medical Devices
Directive 93/42/EEC.



Intelligent Orthopaedics

PO Box 2681
Stoke on Trent
UK ST4 9BE

T +44 (0) 17 82 84 78 40

www.intelligent-orthopaedics.co.uk

Product: External Fixator Bone Pin

Part Number: SIFP - DDDLLLSS

Details of product:

Single use external fixator pin for use with Intelligent
Orthopaedics Fracture Management Devices (t > 30 days)

DDD = diameter

LLL = overall length

SS = style number

EC Product Class Analysis:

Annex IX: Rule 8

Long Term Use surgically invasive device

Class IIb

Notified Body

Amtac Medica

Technical Files are held at Intelligent Orthopaedics Ltd.

A Complaints and Vigilance Procedure is in action and details are supplied
with the purchase contract, they may also be obtained from <http://www.intelligent-orthopaedics.co.uk>.

Approved by: *C Ian Moorcroft*

Date: *4/1/06*

Dr C Ian Moorcroft
Technical Director
Intelligent Orthopaedics Ltd.

QMF_TF02

Building 103
Campbell Road
Stoke on Trent
ST4 4DE



Intelligent Orthopaedics

Declaration of Conformity

EC Declaration of Conformity under Annex VII of the Medical Devices Directive 93/42/EEC

Product: IOS external fixator
Part Number: 2-0-0 (rev 1.0 onwards)

Details of product:

Single use external fixator for the treatment of unstable fractures.
Registered with the MHRA as Product Category G7: Orthopaedic casting/support products and accessories

EC Product Class Analysis:

Annex IX, Rule 1. Non invasive, Class 1

Technical Files are held at Intelligent Orthopaedics Ltd.

A Complaints and Vigilance Procedure is in action and details are supplied with the purchase contract, they may also be obtained from <http://www.intelligent-orthopaedics.co.uk>.

Approved by:

Date:

18/4/05

Dr C Ian Moorcroft
Technical Director
Intelligent Orthopaedics Ltd.

Intelligent Orthopaedics Ltd
Company Registration No: 4504094 Registered Office: Glebe Court, Stoke on Trent, ST4 1ET

2-0-0 QMF_TF02

Rev:1.1
Date 15/03.2005

Approved by: C.I.Moorcroft
Page 1 of 1